

DDS CALORIMETERS

Scientific Analytical Calorimeter Solutions



CAL3K-S CALORIMETER

Oxygen Bomb Calorimeter

MANUFACTURING SUPERB CALORIMETERS FOR TODAY'S ANALYTICAL NEEDS www.ddscalorimeters.com

CAL3K-S CALORIMETER

CAL3K-S is the fifth in the range of innovative new oxygen bomb calorimeters from DDA Calorimeters. The new range, from the engineers who designed the CAL2K Oxygen Bomb Calorimeter range, is a totally dry type calorimeter, making it environmentally friendly and energy efficient. The CAL3K-S also boasts the smallest carbon footprint in its class.

The CAL3K-S is an economical system and is typically used in applications where low sample analysis is required, like in Food/Feed Analysis, Alternative Energy, Manufacturing, Coal and Oil Production, Research, Universities, and Quality Assurance. In short: wherever the calorific value of a solid or liquid sample must be determined, the CAL3K-S is the economical calorimeter of choice.

COMPLETE SYSTEM

The CAL3K-S system includes a calorimeter, 1 x standard thread bomb vessel, and oxygen filling station. A spare bomb vessel (3K-5S) can be purchased, however this will not increase the speed of the calorimeter system.

The calorimeter system can complete 1 sample per hour. Part of the CAL3K-S system is the external oxygen filling station (3K-3), standard across all manual filling calorimeter systems.



COMPLETE SYSTEM For use with the cal3k-s.

The CAL3K-S is your entry level system suitable for low throughput without compromising on accuracy and repeatability. It is ideal for use in Food and Nutrition, Animal Feed Analysis, Universities and Research Departments, Alternative Energy Research, Waste and Waste Product Analysis.

The following are included:

- 3K-3 Manual Oxygen Filling Station
- 3K-5S Standard Bomb Vessel

The following can be optionally added to the system:

- Analytical Balance (sold separately)
- High Pressure Oxygen Regulator



The complete CAL3K-S oxygen bomb calorimeter system contains all the parts and consumables necessary to set up the system. The installation kits included with the setup of the calorimeter contain consumables for approximately 200 samples, depending on the type of sample being analysed, such as coal, food, or feed samples. Other samples, such as oil, might use more consumables as they are corrosive and could cause wear and tear.

Additional consumables can be purchased separately from DDS or an authorised agent.

The sample repeat speed is 30 minutes. With Built-In Fan Cooler.

The vessel is manually filled with oxygen via the external manual oxygen filling station (3K-3).

1 Vessel system. Not expandable.

The system can analyse 1 sample per hour; a total of 8 samples per day.

CAL3K-S CALORIMETER

ADVANCED CAL3K-S FEATURES



TEMPERATURE CONTROL No temperature control of room/lab required

16 CALIBRATION FIELDS For different mode and different calorimeters



FAULT FINDING Extensive fault finding and testing

Thread bomb vessel

STANDARD THREAD BOMB VESSEL



STEP-BY-STEP HELP

Screen prompts assist with step-by-step instructions to operate the calorimeter



USER FRIENDLY User Friendly Operation

10 CALIBRATION AVERAGE For variable amount of calibration average to suit your application

LOW POWER CONSUMPTION Very low power consumption. No temperature controlling required.

ECO FRIENDLY Eco Friendly - small carbon footprint. No water, low power consumption.

TEMPERATURE RANGE Extensive temperature range from 0°C to 70°C.

EVENT LOGGING Built-in event logging for ~6000 events





COMPENSATION* Compensation for firing energy and sulphur

PRESET FIELDS One default setting per mode

OPERATING PARAMETERS Operating parameters can be changed via USB interface in experimental mode

RESTRICT ACCESS Operating parameter access is password restricted

LARGE STORAGE Up to 1024 results storage

INTELLIGENT VESSEL Intelligent vessel with built-in temperature sensing

LINEAR SENSORS Linear temperature sensing with platinum sensors



SAFETY Safety checks guarantee the safety of the operator.



OPTIONAL BALANCE INTERFACE Balance interface with baud speed setting

C

IMPROVED INITIAL TIMING Based on drift and time, or time after drift

LIMS

FULL LIMS SUPPORT For the assignment, scheduling, and tracking of samples.



FILTER DATA ON EVENTS Extensive testing and detailed data viewing





NO WATER REQUIRED No Water Bucket. No Spillage. No Measuring.



MANUAL OXYGEN FILLING Makes use of an external oxygen filling station

CAL3K-S CALORIMETER

The CAL3K Bomb Calorimeter Installation 3K-S-Kit includes :

- Power Supply (External 12V/1.25A)(3K-1-055)
- PC Keyboard PS2 (3K-1-061)
- Preparation Stand (3K-4-49)
- Tweezers (3K-1-081)
- Certified Benzoic Acid Tablets (150 x 0.5g tablets per bottle)(3K-4-71)
- Wire Brush (3K-4-106)
- Printed Installation Manual (INSTALLATION-MANUAL)
- USB 32Gb Green Memory Flash Drive (3K-1-043)

CAL3K-S KIT AND CONSUMABLES

The CAL3K Standard Thread Vessel Installation 3K-5S-Kit includes :


- Complete Top and Bottom Centre Electrode (3K-4-96)
- Outside Electrode (3K-4-37)
- Single Crucibles (3K-4-47)
- Deflector Plate (3K-4-92)
- Firing Cotton (Bundle of 100 threads)(3K-4-65)
- Firing Wire (5 per packet)(3K-4-93)
- Lid O-Ring (3K-4-94)
- Top and Bottom O-Ring in Vessel Lid (3K-4-22)

The CAL3K Filling Station Installation 3K-3-Kit includes :

- Nozzle O-Ring for Defiller Cap and Filling Station (3K-3-29)
- Nozzle O-Ring Used in 3K-3 Jet Assembly (3K-1-080)
- Oxygen Regulator Connection Kit (3K-3-21)
- Defiller Cap (3K-3-22)
- Emergency Deflate Cap (3K-3-18)
- High Pressure Pipe 4mm (Clear/White) (PIPE-4MM-CLEAR)



TECHNICAL SPECIFICATIONS



| Specification | Information |
|------------------------------------|---------------------------|
| Working (Operating) Temperature | 15-70°C |
| Storage Temperature | 0-70°C |
| Temperature Resolution | 0.000001°C |
| Reproducibility/Repeatability | 0.2% RSD |
| Resolution | 0.0001 MJ/Kg |
| Results per hour | 1 Sample per hour |
| Measuring range max. | 99MJ, 99000J |
| Working temperature min. | 1°C |
| Working temperature max. | 50°C |
| Temperature Measurement Resolution | 10ppm (parts per million) |
| Cooling Medium | Fan Air |

TECHNICAL SPECIFICATIONS

| Specification | Information | |
|----------------------------------|--------------------------------------|--|
| Type of Cooling | Built-in cooling | |
| Oxygen Operating Pressure Max | 40 bar | |
| Balance/Scale Interface | RS232, 1200 to 38400 Baud (settable) | |
| Printer Interface | RS232, 1.2Kb to 115.2Kb | |
| Power Input | 2.4W | |
| Interface External Keyboard | PS2 | |
| Oxygen Filling | Manual | |
| De gasification | Manual | |
| Halogen (Decomposition) Vessel | Yes, optional | |
| Analysis according to DIN 51900 | Yes | |
| Analysis according to ASTM D240 | Yes | |
| Analysis according to ASTM D4809 | Yes | |
| Analysis according to ASTM D5865 | Yes | |
| Analysis according to ASTM E711 | Yes | |
| Analysis according to ISO 1928 | Yes | |
| Permissible Ambient Temperature | 1-35°C | |
| Permissible Relative Humidity | 80% | |
| RS232 Interface | Yes | |
| Voltage | 220-240 / 100-120V, 12VDC, 1Amp | |
| Frequency | 50/60 Hz | |

Please Note : Technical Specifications subject to change without prior notice.

Please contact our team for accurate technical specifications at the time.



SYSTEM COMPARISON

| FEATURE | CAL3K-AP | CAL3K-A | CAL3K-S | CAL3K-F |
|----------------------------------|----------------------------------|----------------------------------|----------------------------------|----------------------------------|
| LIMS | 2 x via RS232, or Bluetooth | Yes | Yes | Yes |
| MASS HEAP | No | No | No | No |
| BALANCE INTERFACE | Yes, from 1.2 to 38.4KB |
| RESULT MEMORY | 1024 records, 262KB | 1024 records | 300 records | 300 records |
| TEMPERATURE RESOLUTION | 0.000'001°C | 0.000'001°C | 0.000'001°C | 0.000'001°C |
| DISPLAY | 4 x 40 character LCD | 4 x 40 character LCD | 2 x 20 character LCD | 4 x 40 character LCD |
| KEYBOARD | QWERTY, External, PS2 | QWERTY, External, PS2 | QWERTY, External, PS2 | QWERTY, External, PS2 |
| SAMPLE ID | 16 characters, auto-increment | 16 characters, auto-increment | 16 characters, auto-increment | 16 characters, auto-increment |
| GROUP ID | 16 characters | 16 characters | - | 16 characters |
| VESSEL RECORD | Yes, unlimited | Yes, unlimited | Limited | Limited |
| REAL TIME | Yes | Yes | Yes | Yes |
| CHASSIS IDENTIFICATION | Yes, number | Yes, number | Yes | Yes |
| CALIBRATION | Normal average & PC |
| HISTORY CALIBRATION | Yes, up to 10 runs | Yes, up to 10 runs | Yes, up to 16 runs | Yes, up to 10 runs |
| UNITS | KJ, BTU, CAL | KJ, BTU, CAL | KJ, BTU, CAL | KJ, BTU, CAL |
| RESULT COMPENSATION | Via PC (IntelCal), Default | Via PC (IntelCal), Default | Yes | Yes |
| RESULT VALIDATION | Yes | Yes | Yes | Yes |
| VESSEL PRESS. MONITOR | Up to 100 bar | No | No | No |
| OXYGEN FILLING | Internal, automatic filling | External manual filling station | External manual filling station | External manual filling station |
| DE-FILLING | Automatic | Manual | Manual | Manual |
| MAX CHASSIS RECORDING | Yes | Yes | Yes | Yes |
| CHASSIS NAME | 20 characters, Bluetooth name | 20 characters, Bluetooth name | - | - |
| PASSWORD | CAL3K & PC Password | CAL3K & PC Password | CAL3K & PC Password | CAL3K & PC Password |
| NETWORK MULTIPLE CALORIMETERS | No | No | No | No |
| VESSEL LEAK MONITOR | Yes, flags result and warning | No | No | No |
| EVENT STORAGE | ~6000 events | ~6000 events | - | ~3000 events |
| EVENT TYPES | ~70 different events | ~70 different events | - | ~80 different events |
| EVENT CLASSIFICATION | Operational & Technical | Operational & Technical | - | Yes |

SYSTEM COMPARISON

| FEATURE | CAL3K-AP | CAL3K-A | CAL3K-S | CAL3K-F |
|--|--------------------------------|--------------------------------|----------------------------------|--------------------------|
| VESSEL LOCKOUT, LOCK-IN | Yes, Manual/Auto Linking | Yes, Manual/Auto Linking | Yes, Manual/Auto Linking | Yes, Manual/Auto Linking |
| SAMPLE REPEAT SPEED (FIRST 2 SAMPLES) | 6 min | 4-5 min | 30 min | 7-8 min |
| OPERATOR TIME PER TEST | +/- 3 min | +/- 3 min | +/- 3 min | +/- 3 min |
| COOLING | Air | Air | Built-in | Air |
| COOLING MODES | Ambient/Fixed | Ambient/Fixed | Ambient/Fixed | Ambient/Fixed |
| RSD | <0.1 | 0.1 | 0.2 | 0.1 |
| POWER CONSUMPTION | 0-264 VAC 12W | 0-264 VAC 12W | 12 VAC 6W | 12 VAC 6W |
| POWER SUPPLY | External 12V | External 12V | External 12V | External 12V |
| WATER CONSUMPTION | None | None | None | None |
| REPEATABILITY | <0.1% | 0.1% | 0.2% | 0.1% |
| CALORIMETER TYPE | Dynamic, Isothermal, Adiabatic | Dynamic, Isothermal, Adiabatic | Dynamic | Dynamic |
| NUMBER OF VESSELS | Unlimited | Unlimited | Limited (1) | Unlimited |
| CLOSURE TYPE | Bayonet Lid | Bayonet Lid | Screw Cap (thread) | Bayonet Lid |
| TESTS P/H WITH 2 VESSELS | 10+ | 10+ | 1-2 Sample with 1 Vessel Only | 8+ |
| BOMB VESSEL TYPE | Removable | Removable | Removable | Removable |
| OXYGEN FILLING | Automatic | Manual | Manual | Manual |
| BOMB VESSEL WASHING | Manual | Manual | Manual | Manual |
| PRINTER CONNECTION | RS232 | RS232 | RS232 | RS232 |
| BALANCE CONNECTION | RS232 | RS232 | RS232 | RS232 |
| ENVIRONMENTAL | 5-40°C | 5-40°C | 5-40°C | 5-40°C |
| PRINTING OF RESULTS | Via PC Software | Via PC Software | Via PC or RS2232 | Via PC or RS2232 |
| PC SOFTWARE | Advanced | Advanced | Advanced | Advanced |
| CORRECTION FACTORS | 8 | 8 | 4 | 8 |
| MASS ENTRY | Auto & Manual | Auto & Manual | Auto & Manual | Auto & Manual |
| CE/TUV CERTIFICATE | Yes (Pending) | Yes (Pending) | Yes (Pending) | Yes (Pending) |
| VESSEL DETERMINATIONS | Unlimited | Unlimited | 5000 | Unlimited |
| SPIKING | Yes | Yes | Yes | Yes |
| SELF TESTING | Yes | Yes | Yes | Yes |

SYSTEM COMPARISON

| FEATURE | CAL3K-AP | CAL3K-A | CAL3K-S | CAL3K-F |
|-------------------------|--|--|-------------------------------|---|
| CONNECTIVITY | USB 2.0, 2 x RS232 at 115.2KB for Bluetooth | USB 2.0, 2 x RS232 at 115.2KB for Bluetooth | USB 2.0, 2 x RS232 at 115.2KB | USB 2.0, 2 x RS232 at 115.2KB |
| STATS | Yes | Yes | Yes | Yes |
| PRINTING | Yes, D1 port, 1.2 to 115.2KB | Yes, D1 port, 1.2 to 115.2KB | Yes | Yes |
| MOISTURE COMPENSATION | Yes | Yes | Yes | Yes |
| FOOD FIBRE COMPENSATION | Yes | Yes | Yes | Yes |
| LIMS | Yes | Yes | Yes | Yes |
| RESULT APPROVAL | Yes, keyboard or PC | Yes, keyboard or PC | Yes, keyboard or PC | Yes, keyboard or PC |
| REAL TIME PRINTOUT | Yes, optional customer and parameter header | Yes, optional customer and parameter header | - | Yes, optional customer and parameter header |
| GELATINE CAPSULE | Yes | Yes | Yes | Yes |



The CAL3K-S Oxygen Bomb Calorimeter System can be used with most applications including, but not limited to : Animal Feed Research, University Research, Food/Nutrition Analysis, Explosives Analysis, Coal Analysis, Oil Analysis, and other traditional and non-traditional applications.

For more details and application notes visit our website at www.ddscalorimeters.com

CONTACT US

COMPANY HISTORY

Digital Data Systems (DDS has more than 40 years of experience in calorimetry.

In 1972, DDS produced their first calorimeter, the AMPC (Automatic Micro Processor Calorimeter). The AMPC was a dual water isothermal unit controlled by a microprocessor.

In 1980 work began on a new revolutionary design of vessel, namely the DRY vessel or CP510, which meant that there was no surrounding water jacket. A copper sleeve pressed over the vessel replaced the water jacket and the temperature sensors were placed inside the vessel resulting in the heat transfer being extremely fast. Determination time was significantly reduced, increasing the unit efficiency by 4 times. With the processing power of the microprocessors available at the time, the CP500 Calorimeter was born. The striking "buttercup yellow" colour gave a splash of brightness to the then drab laboratories. In 2002 work began on the CAL2K. The tried and tested DRY system was retained and only the very latest electronic technology was used, including the surface mount devices.

In 2005, DDS came to realize the need for smaller, low volume, inexpensive calorimeter systems, with the same accuracy and reliability of the CAL2K. The ECO was then created as an alternative system to the CAL2K. The ECO is suitable for the following markets: Universities, Research Facilities, Brick Manufacturers, Animal Feed Industries, Food Quality, and Food Production.

In 2007 the new E2K system was developed.

In 2014 work began on the new CAL3K range. The CAL3k-A the first of the range, the CAL3k-F and CAL3K-AP shortly thereafter. In 2018 work began on the new CAL3K-S.

digital data systems (pty) Itd

DDS Calorimeters are proudly manufactured by : Digital Data Systems (Pty) Ltd.

For more information about any of our products visit our website at www.ddscalorimeters.com.

DDS Calorimeters

Email : calo@ddsystems.co.za Tel : +2711 792 9805 www.ddscalorimeter.com

22 Arbeid Avenue Strydom Park Randburg, Gauteng 2196

