LSSD-01 Leak and Seal Strength Detector is professionally applicable to the quantitative determination of seal performance, seal quality, burst pressure, compression resistance, torsion force and joint/disengaging force of flexible packages, aseptic packages, various plastic pilfer-proof closures, flexible tubes, caps and other materials.



Labthin

Professional Technology

- Based on the positive pressure method and controlled by micro-computer, with LCD, menu interface and PVC operation panel
- Dual test methods of restraint distension and unrestraint distension for customer's free choice
- Different test modes of burst, creep, and creep to failure to meet different test requirements
- Optional test range, "one key operation" and other intelligent designs support combinations of non-standard test conditions
- Professional software provides automatic statistics of test data
- Equipped with micro-printer and standard RS232 port for convenient PC connection and data transfer
- Supports LystemTM Lab Data Sharing System for uniform and systematic management of test results and test reports

Test Standards

This test instrument conforms to the following standards:

ISO 11607-1, ISO 11607-2, GB/T 10440, GB 18454, GB 19741, GB 17447, ASTM F1140, ASTM F2054, GB/T 17876, GB/T 10004, BB/T 0025, QB/T 1871, YBB 00252005, YBB 00162002

Applications

Basic Applications	Plastic Composite Bags	Test the compression resistance of various plastic films, aluminum	
		films, paper plastic composite films, aluminum plastic composite	
		films and other packaging bags	
		Including various flexible tubes used in daily chemical products and	
	Flexible Tubes	other industries, e.g. flexible tubes of toothpaste, face cream,	
		cosmetics, pharmaceutics and food	
	Creep Test	Including various packaging bags and boxes	
	Creep to Failure Test	Including various packaging bags and boxes	
Extended Applications	Burst Test of Blister	Including various blister packs	
	Packs		
	Aerosol Valves	Test seal performance of various aerosol valves, e.g. vales of	

LSSD-01 Leak and Seal Strength Detector is applicable to the determination of:

Labthink Instruments Co., Ltd. 144 Wuyingshan Road, Jinan, P.R.China (250031) Phone: +86-531-85068566 FAX: +86-531-85062108 Labthink International, Inc. 200 River's Edge Drive, Medford, MA, 02155, U.S.A. Phone: +1-617-830-2190 FAX: +1-781-219-3638

(ES) Equipements Scientifiques SA - Département Bio-Tests & Industries - 127 rue de Buzenval BP 26 - 92380 Garches Tél. 01 47 95 99 90 - Fax. 01 47 01 16 22 - e-mail: bio@es-france.com - Site Web: www.es-france.com

Labthink®

	pesticide, hair spray, auto spray paint and medical spray packages		
Three-sided Sealing	Test withstand pressurization stress of packaging bags with		
Materials	three-sided seal and one-sided open		
High Pressure Test	The maximum test pressure could reach 1.6 MPa		
	Test seal performance of various pilfer-proof closures, e.g. closures		
Dilfon and of Classes	used in packages for Coke, mineral water, beverage, edible oil, sauce		
Pilfer-proof Closures	(soy, vinegar and cooking wine), three-piece cans (beer and		
	beverage) and paper cans (cylinder shape for potato chips)		

Technical Specifications

Specifications	LSSD-01	
	0 ~ 600 KPa; 0 ~ 87.0 psi (standard)	
Test Range	0 ~ 1.6 MPa; 0 ~ 232.1 psi (optional)	
Cog Injection Head	Φ10 mm (standard)	
Gas Injection Head	$\Phi4$ mm, $\Phi1.6$ mm (optional)	
Gas Supply Pressure	0.4 MPa ~ 0.9 MPa (outside of supply scope)	
Port Size	Φ8 mm PU Tubing	
Instrument Dimension	300 mm (L) x 310 mm (W) x 180 mm (H)	
Pedestal Size	305 mm (L) x 356 mm (W) x 325 mm(H)	
Power Supply	220VAC 50Hz / 120VAC 60Hz	
Net Weight	23 kg	

Configurations

Standard Configurations	Instrument, Test Frame and Micro-printer	
	Professional Software, Communication Cable, Test Accessories (Fixtures for Closed	
Ontional Darta	Package Restraining Test, Open Package Unrestrained Test, Plastic Pilfer Proof	
Optional Parts	Closure test, Flexible Tubes Test and Aerosol Vale Seal Performance), and Test Device	
	for Blister Pack Seal Performance	
Note	1. The gas supply port of the instrument is $\Phi 8 \text{ mm PU}$ tubing;	
THOLE	2. Customers will need to prepare for gas supply.	

Please Note: Labthink is always dedicated to the innovation and improvement of product performance and function. Therefore, technical specifications are subject to change without further notice. Please visit our website at www.labthink.com for the latest updates. Labthink reserves the rights of final interpretation and revision.

 Labthink Instruments Co., Ltd.
 144 Wuyingshan Road, Jinan, P.R.China (250031)
 Phone: +86-531-85068566
 FAX: +86-531-85062108

 Labthink International, Inc.
 200 River's Edge Drive, Medford, MA, 02155, U.S.A.
 Phone: +1-617-830-2190
 FAX: +1-781-219-3638

(ES) Equipements Scientifiques SA - Département Bio-Tests & Industries - 127 rue de Buzenval BP 26 - 92380 Garches Tél. 01 47 95 99 90 - Fax. 01 47 01 16 22 - e-mail: bio@es-france.com - Site Web: www.es-france.com