

# Dry Gluten DG 2020

8

- CE certified.
- Heater and control units are separate.
- Possibility to see the test time on digital display.
- Dries gluten of wheat and flour samples.
- Operating temperature is 150°C.
- Temperature is kept stable by means of a thermostat.
- Teflon coating provides effective drying and easy removal of sample being tested.
- Stops automatically warning with a beep signal.
- 220 V 50 Hz electricity.



**ERKAYA**  
INSTRUMENTS