


# WPF18 Probe

## 1 MHz – 18 GHz



- Electric field measurement
- Isotropic & True RMS measurement
- Excellent attenuation at 50/60 Hz
- Measurements in accordance with International Standards

- 1 MHz – 18 GHz
- E
- RMS
- ISOTROPIC



**Telecommunications**  
Certification and audit of telecommunication services (GSM, 3G, LTE, TDT, FM, WiFi, etc.).



**Industry**  
Assessment of industrial processes for worker's exposure protection.



**Radar/Aerospace**  
Assessment of radar and aerospace sites and personnel exposure protection.



**Labs/R&D**  
RF exposure protection of R&D and labs personnel.



### Technical Specifications

	WPF18	WPF18-FAST
Part number	WWP0903	WWP0913
Frequency range	1 MHz – 18 GHz	
Sensor type	Isotropic RMS diode technology	
Type of frequency response	Flat	
Measurement range	0.5 – 1000 V/m (CW) 0.5 – 50 V/m (True RMS)	
Dynamic range	66 dB	
CW damage Level	1200 V/m	
Integration time	400 ms (higher 50/60 Hz immunity)	4 ms (fast mode switchable)
Frequency response (*)	>-3dB @ 1 MHz ±2 dB (2 MHz - 15 GHz) +1/-3 dB (15 GHz - 18 GHz)	
Linearity	±1 dB (1 – 3 V/m) ±0.5 dB (3 - 200 V/m)	
Isotropic deviation	±1 dB (1 MHz - 12 GHz) ±2 dB (12 GHz - 18 GHz)	
Resolution	< 5 %	
Calibration	ISO 17025 Accredited Calibration (ILAC) (**)	
Calibration period	24 months (recommended)	
Temperature range	-20 °C to 50 °C	
Temperature response	+0.1 / -1 dB (related to 20 °C)	
Dimensions	28,4 cm X 6 cm Ø (11.2" x 2.3" Ø)	
Weight	95 g (3.4 oz.)	

(\*) The frequency response can be corrected with the SMP2/SMP3 by using the correction factors stored in the probe (ISO 17025 accredited calibration).  
(\*\*) In combination with SMP2, SMP3, MonitEM, MonitEM-LAB devices.

Compatible with **SMP3**, **SMP2**, **MonitEM**, **MapEM**, **MonitEM-Lab**



WPF18\_EN\_2601\_v2.3



ES France - Département Tests & Mesures  
127 rue de Buzenval BP 26 - 92380 Garches



Tél. 01 47 95 99 45



e-mail : [tem@es-france.com](mailto:tem@es-france.com)  
Site Web : [www.es-france.com](http://www.es-france.com)