# GPM-8320/8330 Specifications

The specifications apply when warmed up for at least 30 minutes and operates under the slow rate & 18~28 °C.













with GPM-DA12

With GPM-DA12

### Input

Item	Specifications			
Innut tuno	Voltage	Floating input thro	ugh resistive voltage divider	
Input type	Current	Floating input through shunt		
	Voltage	15V, 30V, 60V, 150V, 300V, 600V and 1000V		
	Current			
Measure range	Direct input	0.5A, 1A, 2A, 5A, 10A and 20A		
	Sensor input	EX1: 2.5 V, 5 V, 10 V	V	
		EX2: 50 mV, 100 m	V, 200 mV, 500 mV, 1 V, 2 V	
	Voltage		Input resistance: approach 2 MΩ	
	Current			
Input impedance	Direct input ra	nge 0.5A ~ 20A	Input resistance: approach 5 mΩ	
input impedance	Sensor input			
	Input range 2.5V ~ 10V (EX1)		Input resistance: approach 100 kΩ	
	Input range 50mV ~ 2V (EX2)		Input resistance: approach 20 kΩ	
	Voltage		peak value of 1.5kV or RMS value of 1kV, whichever is less	
Continuous maximum	Current			
allowable input	Direct input range 0.5A ~ 20A		peak value of 100A or RMS value of 30A, whichever is less	
	Sensor input		peak value less than or equal to 5 times of the rated range	
Input bandwidth	DC, 0.1 Hz ~ 100kHz			
Continuous maximum	600 Vrms, CAT Ⅱ			
Common-mode voltage				
Line filter	select OFF or ON (cut off frequency of 500 Hz)			
Frequency filter	select OFF or ON (cut off frequency of 500 Hz)			
	Simultaneous conversion voltage and current inputs			
A/D converter	Resolution 16bits			
	Maximum conversion rate Approx. 300kHz			
Display update interval	When the data update interval is 100 ms the numeric display 10 items display update			
	interval is 200	ms.		







When the data update interval is 100 ms or 250ms and the numeric value display is set to Matrix or ALL Items display update interval is 500 ms. The waveform display update intervals are approximately 1s.

### **Voltage and Current Accuracy**

Item	Specifications			
	Temperature		23 ± 5℃	
	Humidity		30~75% RH	
	Input waveform		Sine wave crest factor = 3	
	common-mode voltage		0 V	
Requirements	Number of displayed digi	ts	5 digits	
	Frequency filter			ltage or current of 200 Hz or less
	After 30 minutes after wa	arm-up ti	me has passed	
	After measurement range	e is chan	ged (zero-level compen	sation)
	Update interval is 250 ms	5		
			of reading + 0.2% of ran	
	0.1 Hz ≤ f < 45 Hz	± (0.1 %	of reading + 0.2 % of ra	ange)
	45 Hz ≤ f ≤ 66 Hz	± (0.1 %	of reading + 0.05 % of r	range)
Accuracy	66 Hz < f ≤ 1 kHz	± (0.1 %	of reading + 0.2 % of ra	ange)
	1 kHz < f ≤ 10 kHz	± (0.07 *f) % of reading + 0.3% of range)		
	10 kHz $<$ f $\le$ 100 kHz $\pm$ (0.5 % of reading + 0.5 % of range) $\pm$ [{0.04x(f-10)}% of reading]			
	Values for voltage in excess of 750V for which 30kHz < f < 100kHz are reference only.			
Temperature coefficient	Add ±0.03% of reading/°C within the range 5 to 18°C or 28 to 40°C.		e range 5 to 18°C or 28 to 40°C.	
When the line filter is	45 ~ 66 Hz Add 0.3 % of reading			
turned ON	< 45 Hz Add 1 % of reading			
	accuracy obtained by doubling the measurement range error for the accuracy when the			
factor is set to 6 or 6A	crest factor is set to 3			
Accuracy changes	When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1			
caused by data update	kHz accuracy.			
interval	A 1 1 0 000/		1.	
Influence of	Add 0.02% of range/°C to			
temperature changes	Add the following value t			500 A /00
after zero-level	0.5 A/1 A/2 A/5 A/10 A/2			500 μA/°C
compensation or range	External current sensor input (/EX1) 1 mV/°C			
change	External current sensor in		<i>'</i>	50 μV/°C
•	•	ibling the	e measurement range e	error for the accuracy when the
factor is set to 6 or 6A		crest factor is set to 3		
Accuracy changes caused by data update	When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1			
interval	kHz accuracy.			
incerval				

#### **Active Power Accuracy**

Item	Specifications			
Paguiroments	same as the conditions for voltage and current.			
Requirements	Power factor	1		
	DC	(0.1 % of reading + 0.2 % of range)		
	0.1Hz ≤ f < 45 Hz	± (0.3 % of reading + 0.2 % of range)		
Accuracy	45 Hz ≤ f ≤ 66 Hz	± (0.1 % of reading + 0.05 % of range)		
	66 Hz < f ≤ 1kHz	± (0.2 % of reading + 0.2 % of range)		
	1 kHz < f ≤ 10 kHz	$\pm$ (0.1 % of reading + 0.3 % of range) $\pm$ [{0.067x(f-1)}% of reading]		
	10 kHz < f ≤ 100 kHz	$\pm$ (0.5 % of reading + 0.5 % of range) $\pm$ [{0.09x(f-10)}% of reading]		
	when power factor (λ) = 0 (S: apparent power)			
Influence of power	± 0.1 % of S for 45 Hz ≤ f ≤ 66 Hz			
factor	± {(0.1 + 0.15 × f) % of S } for up to 100 kHz as reference data			
	•f is frequency of input signal in kHz			



	when $0 < \lambda < 1$ ( $\Phi$ : phase angle of the Voltage and current)		
	(power reading ) × [(power reading error%) + (power range %) × (power range / indicated		
	apparent power value) +	{tanΦ× (influence when λ=o)%}]	
When the line filter is	45 ~ 66 Hz	Add 0.3 % of reading	
turned ON	< 45 Hz	Add 1 % of reading	
Temperature coefficient	same as the temperature	coefficient for voltage and current	
Accuracy when the crest	accuracy obtained by dou	ibling the measurement range error for the accuracy when the	
factor is set to 6 or 6A	crest factor is set to 3		
Accuracy of apparent	voltage accuracy + curren	t accuracy	
power S			
Accuracy of reactive	accuracy of apparent pow	ver + (√1.0004 - λ2) - (√1 - λ2) ×100 %	
power Q			
Accuracy of power	± [(λ-λ/1.0002)+   cosø-c	$\cos\{\phi + \sin - 1 \text{ (influence from the power factor when } \lambda = 0\%/100)\} \ \ ]$	
factor λ	±1 digit when voltage and current are at the measurement range rated input		
Accuracy of phase	±[   ø-cos-1(λ/1.0002)	+ sin-1 (influence from the power factor when $\lambda$ = 0 % / 100)] $\pm$ 1	
difference Φ	digit when voltage and cu	irrent are at the measurement range rated input	
Accuracy when the crest	accuracy obtained by dou	ibling the measurement range error for the accuracy when the	
factor is set to 6 or 6A	crest factor is set to 3		
Accuracy changes	When the data update in	terval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1	
caused by data update	kHz accuracy.		
interval			

## **Voltage, Current and Active Power Measurements**

Item	Specifications			
Measurement method	Digital sampling method			
Crest factor	3 or 6 (6A)			
Wiring system	Single-phase, tw	o-wire (1 P2 W)		
Range select	Select manual or	r auto ranging		
	Auto-range incre	ease		
	The range is upped when any of the following conditions is met.			
	Crest factor 3	Vrms or Irms exceeds 130% of the currently set measurement range.		
		Vpk, Ipk value of the input signal exceeds 300% of the currently set		
		measurement range.		
	Crest factor 6	Vrms or Irms exceeds 130% of the currently set measurement range.		
		Vpk, Ipk value of the input signal exceeds 600% of the currently set		
		measurement range.		
	Crest factor 6A	Vrms or Irms exceeds 260% of the currently set measurement range.		
		Vpk, Ipk value of the input signal exceeds 600% of the currently set		
	measurement range.			
Auto range	Auto-range decline			
	The range is downed when all of the following conditions are met.			
	Crest factor 3	Vrms or Irms is less than or equal to 30% of the measurement range.		
		Vrms or Irms is less than or equal to 125% of the next lower		
		measurement range.		
		Vpk, lpk value of the input signal exceeds 300% of the currently set		
	Crest factor 6 or	measurement range.  6A Vrms or Irms is less than or equal to 30% of the measurement range.		
	Crest factor 6 or	Vrms or Irms is less than or equal to 125% of the next lower		
		measurement range.		
		Vpk, lpk value of the input signal exceeds 600% of the currently set		
		measurement range.		
	Vrms (the true RMS value of voltage and current)			
	VOLTAGE MEAN (the rectified mean value calibrated to the RMS value of the voltage and the			
Display mode Switching	true RMS value of the current)			
. ,	AC			
	DC			



	I <sub>-</sub>			
Measurement	Select voltage, current, or off			
synchronization source	In the case of Auto Update Rate, select the voltage or current from the equipped element.			
Line filter	Select OFF or ON (cutoff frequer	ncy at 500 Hz).		
Peak measurement	Measures the peak (max, min) value of voltage, current or power from the instantaneous			
reak illeasurement	voltage, instantaneous current c	or instantaneous power that is sampled.		
Zero-level compensation	Removes the internal offset of t	he measure unit (After measurement range is changed)		
	Voltage	Vrms , Vmn, Vdc , Vac		
	Current	Irms , Idc , Iac		
	Active Power	P		
	Apparent Power	VA		
	Reactive power	VAR		
	Power Factor	PF		
Measurement	Crest Factor	CFI, CFV		
	Phase Angle	DEG		
parameters	Frequency	IHz and VHz		
	Voltage Peak	V+pk and V-pk		
	Current Peak	I+pk and I-pk		
	Active Power Peak	P+pk and P-pk		
	Total Harmonic Distortion	THDI and THDV		
	Mathematical Computation	MATH		
	Maximum Current Ratio	MCR		

#### **Frequency Measurement**

riequency ivieasureme				
Item	Specifications			
Measurement item	Voltage and current			
	Data update interval	Measurement Frequency Range		
	0.1 s	20 Hz ≤ f ≤ 100 kHz		
	0.25 s	10 Hz ≤ f ≤ 100 kHz		
	0.5 s	5 Hz ≤ f ≤ 100 kHz		
	1 s	2.0 Hz ≤ f ≤ 100 kHz		
	2 s	1.0 Hz ≤ f ≤ 100 kHz		
	5 s	0.5 Hz ≤ f ≤ 100 kHz		
Measurement frequency	10 s	0.2 Hz ≤ f ≤ 100 kHz		
range	20 s	0.1 Hz ≤ f ≤ 100 kHz		
	Auto ( * ) $0.1 \text{ Hz} \le f \le 100 \text{ kHz}$			
	(*) Limit of the measurement lower limit frequency by the Timeout setting			
	Timeout	lower limit frequency		
	1 s	2.0 Hz		
	5 s	0.5 Hz		
	10 s	0.2 Hz		
	20 s	0.1 Hz		
Measurement range	Auto switching among six types: 100mHz, 1 Hz, 10 Hz, 100 Hz, 1 kHz, 10 kHz, and 100 kHz.			
Frequency filter	Select OFF or ON (cut off frequency of 500 Hz)			
	Requirements	When the input signal level is 30% or more of the measurement		
Accuracy		range If the crest factor is set to 3.		
		(60% or more if the crest factor is set to 6 or 6A)		
		• Frequency filter is ON when measuring voltage or current of 200 Hz		
		or less.		
	± (0.06% of reading)			

#### Integration

Item	Specifications
Mode	Select manual integration mode, standard integration mode, or repetitive integration mode.
Timer	Automatically stop integration by setting a timer.
	Selectable range: 0 hours 00 minutes 00 seconds to 9999 hours 59 minutes 59 seconds
Accuracy	±(Power accuracy (or current accuracy) + 0.1% of reading) (fixed range)





Range setting	Auto range or fixed range is available for Integration
Timer accuracy	±0.02%
Remote control	Start, stop and reset operations are available using an external remote signal. (option)

#### **Harmonic Measurement**

authorite Wedsareniere				
Item	Specifications			
Measured item	Voltage, Current, Power			
Measured method	Zero-cross simultaneo	us calculation metho	od	
Frequency range	10 Hz to 1.2 kHz.			
EET data longth	4096			
FFT data length	(Auto switch when bo	th 50Hz/60Hz and up	odate rate must be gi	reater than or equal to 0.5s)
Sample rate, window	Fundamental	Sample rate	Window Width	upper limit of Analysis
width, and upper limit	Frequency			orders
of Analysis orders*	45 Hz to 55 Hz	f × 512	10	50
of Affaiysis ofders	54 Hz to 66 Hz	f x 512	12	50
FFT data length	1024	1024		
	Fundamental	Sample rate	Window Width	upper limit of Analysis
	Frequency			orders
Sample rate, window	10 Hz to 67 Hz	f × 1024	1	50
width, and upper limit of Analysis orders*	67 Hz to 150 Hz	f × 512	2	32
	150 Hz to 300 Hz	f × 256	4	16
	300 Hz to 600 Hz	f × 128	8	8
	600 Hz to 1200 Hz	f × 64	16	4
	Frequency	Voltage	Current	Power
	10 Hz ≤ f < 45 Hz	0.15% of reading	0.15% of reading	0.35% of reading
Accuracy		+ 0.35% of range	+ 0.35% of range	+ 0.50% of range
	45 Hz ≤ f < 440 Hz	0.15% of reading	0.15% of reading	0.25% of reading
		+ 0.35% of range	+ 0.35% of range	+ 0.50% of range
	440 Hz ≤ f < 1.2kHz	0.20% of reading	0.20% of reading	0.40% of reading
		+ 0.35% of range	+ 0.35% of range	+ 0.50% of range
* FOUL /COLL C 1: 1.1				

<sup>\* 50</sup>Hz/60Hz Compliant IEC61000-4-7 (update rate must be > 0.5s)

#### **D/A Output (Options)**

Item	Specifications
Output voltage	±5 V FS (approach ±7.5 V maximum) against each rated value.
Number of output	12
channels	
Output items	Set for each channel : V, I, P, VA, VAR, PF, DEG, VHZ, IHZ, Vpk, Ipk, WP, WP±, q, q±, Off
Accuracy	±(accuracy of each measurement item + 0.2% of FS)(FS = 5 V)
D/A conversion	16 bits
resolution	
Minimum load	100 kΩ
	Same as the data update interval.
Update Interval	In the case of Auto Update Rate, update interval is equal to signal interval. More than
	100ms.
Temperature coefficient	±0.05%/°C of FS

## **Remote Control Input/Output Signal (Options)**

ltam	Specifications
Item	specifications
Remote control input	EXT HOLD, EXT TRIG, EXT START, EXT STOP, EXT RESET
signal	
Remote control output	INTEG BUSY
signal	
I/O level	TTL
I/O logic format	Negative logic, Falling edge





<sup>\*</sup> Harmonic calculation: FFT method in which FFT data length is divided into 2 types: 1024 and 4096.

<sup>\*</sup> FFT data length automatically switches in accord with the Frequency and Update Rate of measured signal.



#### **Digital IO Signal (Options)**

Item	Specifications
I/O control output signal	OUT1, OUT2, OUT3, OUT4
I/O level	TTL
I/O sink current	Max 100mA (per/ch)

<sup>\*</sup> Q (VAR), S (VA), λ (PF) and Φ (DEG) are originated from the measured values including voltage, current and active power which go through computation process. In respect to distorted signal input, accordingly, the value acquired from other instruments, which employ different methods, may differ from that acquired from GPM-8310 unit.

#### General

Display	5" TFT LCD
Interfaces	RS-232C, USB host/device, LAN
Power Source	AC 100-240V, 50-60Hz
Power Consumption	35VA max.
Dimensions & Weight	220(W) x 132(H) x 402.5(D) mm (w/t bumpers), Approx. 3.85kg





<sup>\* &</sup>quot;Zero" will be shown for S or Q and "--" will be displayed for λ and Φ when either current or voltage is less than 0.5% of the rated range (less than or equivalent to 1% when crest factor is set 6).