



PORTABLE GAS ANALYSER | LANDFILL & CONTAMINATED LAND

The Geotech GA5000 is a landfill and contaminated land portable gas analyser, with available gas measurements of CH_4 , CO_2 , O_2 , H_2S and CO. It is easy to use and calibrate, benefiting from our market leading reliability and helping you to standardise monitoring routines, whilst supporting environmental legislation compliance.



SECTOR

Landfill Land remediation

APPLICATIONS

- Landfill gas monitoring
- Waste to energy
- Site investigation
- Contaminated land







FEATURES

- Certified: ATEX, IECEx, CSA, MCERTS and UKAS calibration (ISO17025)
- Measures % CH₄, CO₂, and O₂
- Measures barometric pressure and relative pressure
- Peak and previous readings shown
- Choice of user settings and simple gas reading function
- Simultaneous display of all gases
- 3 year warranty
- Modular and upgradeable
- Memory: 2,000 IDs* and 4,000 readings (* with GAM software)
- Data logging and profiling function
- Up to 6 gases monitored

BENEFITS

- Easy to use and calibrate
- Supports environmental legislation compliance
- Market leading reliability
- Standardises monitoring routines
- Easy transfer of data

OPTIONS (AVAILABLE AT PURCHASE OR LATER)

- Choice of additional gases including:
 H₂S to 10,000ppm
 - H₂ compensated CO up to 10,000ppm (1%)
- Borehole gas flow (I/h)
- Flow logging for improved borehole analysis
- GPS / field navigator
- Gas Analyser Manager software for data download
- ATEX certified anemometer 0-40 m/s

© Product designs and specifications are subject to change without notice. User is responsible for determining suitability of product.

Geotech

geotechuk.com
@ sales@geotech.co.uk
+44 (0)1926 338111

(ES) Equipements Scientifiques SA - Département Bio-Tests & Industries - 127 rue de Buzenval BP 26 - 92380 Garches Tél. 01 47 95 99 90 - Fax. 01 47 01 16 22 - e-mail: bio@es-france.com - Site Web: www.es-france.com

GA5000 TECHNICAL SPECIFICATIONS



POWER SUPPLY				
Battery type	Rechargeable nickel met	tal hydride battery pack (n	ot user replaceable)	
Battery life	Typical use 8 hours from fully charged			
Battery charger	Separate intelligent battery charger powered from mains supply (100-240V)			
Charge time	Approximately 4 hours from complete discharge			
GAS RANGES	, ,			
	CO and CU	Du dual wavelength infr	and concer with reference	channel
Gases measured	CO ₂ and CH ₄	By dual wavelength infrared sensor with reference channelBy internal electrochemical sensor		
	O_2 CO (H ₂ compensated), H ₂ S, NH ₃ and H ₂ (optional)	· ·		
	A full range of internal gas cells can be specified at the time of manufacture			
Standard gas cells	Cell	Range	Typical accuracy (range : accuracy)	Typical accuracy (range : accuracy)
	CH ₄	0-100%	0-70% : ±0.5% (vol)	70-100% : ±1.5% (vol)
	CO ₂	0-100%	0-60% : ±0.5% (vol)	60-100% : ±1.5% (vol)
	02	0-25%	0-25% : ±1.0% (vol)	
	Cell	Range	Typical accuracy	
	СО	0-500ppm	±2.0% FS	
	СО	0-1,000ppm	±2.0% FS	
	СО	0-2,000ppm	±2.0% FS	
	CO (H ₂) *	0-2,000ppm	±1.0% FS	
Optional gas cells	H ₂ S	0-50ppm	±1.5% FS	
	H ₂ S	0-200ppm	±2.0% FS	
	H ₂ S	0-500ppm	±2.0% FS	
	H ₂ S	0-1,000ppm	±2.0% FS	
	H ₂ S	0-5,000ppm	±2.0% FS	
	H ₂ S	0-10,000ppm	±5.0% FS	
	NH ₃	0-1,000ppm	±10.0% FS	
	H ₂	0-1,000ppm	±2.5% FS	
Typical accuracies	All typical accuracies quoted are after calibration			
*Hydrogen compensated carbon monoxide measurement	Hydrogen cross gas effect on carbon monoxide approximately 1% Do not use where hydrogen is in excess of 10,000ppm			
Response time, T90	CH_4 ≤ 10 seconds			
	CO ₂	≤10 seconds		
	0 ₂	≤20 seconds		
	СО	≤30 seconds		
	H ₂ S	≤30 seconds		
	NH ₃	≤90 seconds		
	H ₂	≤90 seconds		
PUMP				
Flow	550 ml / min typically			
Flow fail point	-200 mbar vacuum - user settable			
Maximum vacuum restart	-375 mbar approximately with flow rate of approx 80ml / min			

© Product designs and specifications are subject to change without notice. User is responsible for determining suitability of product.

Geotech

🗑 geotechuk.com 🙋 sales@geotech.co.uk 🔲 +44 (0)1926 338111

(ES) Equipements Scientifiques SA - Département Bio-Tests & Industries - 127 rue de Buzenval BP 26 - 92380 Garches Tél. 01 47 95 99 90 - Fax. 01 47 01 16 22 - e-mail: bio@es-france.com - Site Web: www.es-france.com

GA5000 TECHNICAL SPECIFICATIONS CONTINUED



FACILITIES			
Temperature measurement	-10°C to +75°C with optional probe		
Temperature accuracy	±0.5°C with optional probe		
Flow from borehole	0-20 l/hr internal measurement (optional)		
Flow from borehole accuracy	±0.3 l/hr		
Alarm	User selectable alarm levels		
Communications	Via USB lead or wireless Bluetooth**		
Relative pressure measurement	±500 mbar		
Relative pressure accuracy	±4 mbar typically (should be zeroed before reading) to ±15 mbar max		
Barometric pressure measurement	500 to 1500 mbar, ±5 mbar accuracy		
GPS sensor	Location and positioning (optional)		
Available memory	2,000 IDs, 4000 readings**		
ENVIRONMENTAL CONDITIONS			
Operating temperature range	-10°C to +50°C		
Atmospheric pressure range	700 to 1200 mbar		
Relative humidity	0-95% non condensing		
Case seal	IP65		
PHYSICAL			
Weight	1.6kg		
Size	L 220mm, W 155mm, D 60mm		
Case material	ABS / polypropylene with rubber over-moulding		
Keys	Alpha-numeric keypad with "tactile" membrane		
Display	Ultra-clear high resolution 4.3" full colour TFT		
Connections	Colour coded gas inlet, outlet and pressure ports Waterproof USB port, anemometer and charger / temperature probe connections		
Gas sample filters	External user changeable 2.0µm ptfe water traps		
CERTIFICATION RATING			
	II 2G Ex ib IIA T1 Gb (Ta = -10°C to +50°C)		
CERTIFICATION RATING	II 2G Ex ib IIA T1 Gb (Ta = -10°C to +50°C) MC130238		
CERTIFICATION RATING ATEX marking			
CERTIFICATION RATING ATEX marking MCERTS	MC130238		
CERTIFICATION RATING ATEX marking MCERTS ISO17025	MC130238 Calibration to UKAS certificate number 4533 Ex ib IIA T1 (Ta= -10°C to +50°C) (Canada), AEx ib IIA T1 (Ta= -10°C to +50°C) (USA)		

We do, however, reserve the right to change the specification without prior notice as a result of continuing development.



© Product designs and specifications are subject to change without notice. User is responsible for determining suitability of product.

Geotech

🍘 geotechuk.com 🙋 sales@geotech.co.uk 🔲 +44 (0)1926 338111

(ES) Equipements Scientifiques SA - Département Bio-Tests & Industries - 127 rue de Buzenval BP 26 - 92380 Garches Tél. 01 47 95 99 90 - Fax. 01 47 01 16 22 - e-mail: bio@es-france.com - Site Web: www.es-france.com